



THIS DRAWING IS THE PROPERTY OF PRIMARY FLUID SYSTEMS INC. FROM WHICH A WRITTEN CONSENT FOR APPROVAL SHALL BE REQUIRED PRIOR TO COPYING.

| | NO. | NAME | DESCRIPTION |
|---------------------|-----|---|-------------------------------|
| | 1 | MAX ALLOWABLE OPERATING TEMPERATURE | 140°F (60°C) |
| | 2 | MAX ALLOWABLE OPERATING PRESSURE | 15 PSIG |
| | 3 | LEAK TEST | YES |
| CONNECTIONS | 4 | BOTTOM PORT CONNECTION | 1" |
| | 5 | TOP PORT CONNECTION | 1" |
| | 6 | TYPE | NPT |
| MATERIALS | 7 | CYLINDER | GLASS |
| | 8 | END CAPS | SS316L (OPTION ALLOY20, C276) |
| | 9 | GLASS SEALING 'O' RINGS | VITON (OPTION EPDM, PTFE) |
| MATERIAL NON WETTED | 10 | OUTER SHIELD | POLYCARBONATE |
| | 11 | SHIELD SEALING 'O' RINGS | BUNA-N |
| INPUT DATA | 12 | VOLUME SIZE | 2000 ML |
| SCALE | 13 | GRADUATION | 20 ML |
| COMMENTS | 14 | ALL CYLINDERS ARE RATED FOR LIQUID SERVICE ONLY | |
| | 15 | CYLINDERS ARE NOT PRESSURE VESSELS | |




ACCUDRAW®
ACS#2 GLASS CALIBRATION CYLINDER



| REV | DESCRIPTION | DATE | NAME |
|-----|---|-----------|------|
| 1 | END CAP MAT'L SS316L WAS SS316; MAT'L OPTIONS ADDED | 6/24/2021 | MK |

| | | | | |
|---|--|-----------------------|---------------------------|---|
| PRIMARY FLUID SYSTEMS INC. 1050 COOKE BLVD, BURLINGTON, ONTARIO L7T 4A8 | | PROJECT: ACS | SHT: 1 OF 1 | DRAWN BY: NG |
| TITLE: 2000 ML CALIBRATION CYLINDER WITH 1" NPT CONNECTION | | DATE DRAWN: 5/20/2016 | SCALE: 1 : 4 | UNITS: INCH [mm] |
| DWG NAME: ACS CYLINDER DWG TYPE: PLANAR VIEW | | MATERIAL: SEE CHART | DWG NO: ACS#2-2000-GSV_R1 | REV# 1 |
| | | APPROVED BY: | CHECKED BY: | DIMENSIONAL TOLERANCE 0.X = ± 0.25" 0.XX = ± 0.25" 0.XXX = ± 0.25" |



11 X 17
SHEET
SIZE:
B